

Manual Supplement

Manual Title:	187/189 Calibration	Supplement Issue:	1
Part Number:	1584337	Issue Date:	9/03
Print Date:	November 2000	Page Count:	1
Revision/Date:	1, 6/02		

This supplement contains information necessary to ensure the accuracy of the above manual. Enter the corrections in the manual if either one of the following conditions exist:

1. The revision letter stamped on the indicated PCA is equal to or higher than that given with each change.
2. No revision letter is indicated at the beginning of the change.

Change #1

On page 17, Table 3, replace the entry for 350.0°C Input with:

Input	Lower Limits	Upper Limits
350.0 °C	345.5 °C	354.5 °C